



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/511,695

07/01/2005

Attila Herczeg

PB0224

9144

22840 7590 01/12/2009
GE HEALTHCARE BIO-SCIENCES CORP.
PATENT DEPARTMENT
800 CENTENNIAL AVENUE
PISCATAWAY, NJ 08855

EXAMINER

MENON, KRISHNAN S

ART UNIT

PAPER NUMBER

1797

MAIL DATE

DELIVERY MODE

01/12/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/511,695	Applicant(s) HERCZEG, ATTILA	
	Examiner Krishnan S. Menon	Art Unit 1797	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on 05 December 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 2,4,6,7 and 10-12 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 2,4,6,7 and 10-12 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)


- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claims 2,4,6,7 and 10-12 are pending as amended 12/5/08 in the RCE of 8/20/08; claim 12 is independent.

Claim Rejections - 35 USC § 102


1. Claims 2 and 12 are rejected under 35 U.S.C. 102(a/b) as being anticipated by Schmidt et al (US 6,916,420), OR Rogemont (US 4,701,234), OR Kneifel et al (US 4,303,493) OR Applicant's own admission, figures 1,3 and 4 of the disclosure; Claim 12 is anticipated by Grummert, et al (US 6,368,505).



Please note that the US Patents to Grummert and Schmidt cited herein have PCT publications with earlier dates. These references teach the claimed invention – see the figures. They all have the feed, retentate and permeate passages sealed by resins that extend into the holes as well as to the periphery of the holes.

Pictures from applicant's disclosure, disclosed prior art, Grummert, Schmidt and Rogemont (in that order, from left to right) showing the permeate screen. All of them show the resin layer around the feed/retentate hole penetrating the screen material. They do not appear to form any non-uniformities in fluid flow therethrough. Particularly, compare the similarity of applicant's figure 13 to that of Rogemont. The resin layer is taught as penetrating across the screen layer in all the cited references.

The Kneifel reference applicant submitted in an IDS also teaches the same – see the picture captured from the reference below:



Regarding the non-permeable films and the housing, these are parts present in all the prior arts, as admitted by the applicant (see disclosed fig 1). All the references cited teach improvements over the prior arts specific to the feed and/or filtrate screens and/or to the feed, retentate and filtrate apertures and the perimeters of the screens.

A contiguous border is formed at least in Rogemont – (resin layers 3 and 4 in figure 3 – see column 3, line 60 – column 4, line 15), and penetrate the screen evenly – figure 5. Schmidt teaches the same in column 1, lines 16-36. The cross-hatched

perimetrical region on Schmidt figure shown above is a solid frame for the screen.

Applicant's disclosure of the prior art also shows this contiguous resin border with the perimeter of the holes.

2. Claims 4 is rejected under 35 U.S.C. 102(a/b) as being anticipated by, or in the alternative, under 35 USC 103(a) as being obvious over, Grummert, et al (US 6,368,505) OR Schmidt et al (US 6,916,420) OR Rogemont (US 4,701,234) OR Kneifel, OR Applicant's own admission, figures 1,3 and 4 of the disclosure.

Claim 4 adds limitations pertaining to the method of making the cassette, which is not patentable - "[E]ven though product-by-process claims are limited by and defined by the process, determination of patentability is based on the product itself. The patentability of a product does not depend on its method of production. If the product in the product-by-process claim is the same as or obvious from a product of the prior art, the claim is unpatentable even though the prior product was made by a different process." In re *Thorpe*, 777 F.2d 695, 698, 227 USPQ 964, 966 (Fed. Cir. 1985).

3. Claims 6,7,10 and 11 are rejected under 35 U.S.C. 103(a) as being unpatentable over Grummert, et al (US 6,368,505) OR Schmidt et al (US 6,916,420) OR Rogemont (US 4,701,234), OR Kneifel, OR Applicant's own admission, figures 1,3 and 4 of the disclosure.

These claims recite the shape of the apertures with respect to their symmetry about several axes. The references do not specify any particular shape for the

apertures except that they are shown circular. Kneifel does have some symmetry as claimed. However, a difference in the shape of the apertures without any showing of secondary evidence of patentability, is not considered as a patentable difference. Changes of size, shape, etc without special functional significance are not patentable. *Research Corp. v. Nasco Industries, Inc.*, 501 F.2d 358; 182 USPQ 449 (CA 7), cert. denied 184 USPQ 193; 43 USLW 3359 (1974) In *Gardner v. TEC Systems, Inc.*, 725 F.2d 1338, 220 USPQ 777 (Fed. Cir. 1984), cert. denied, 469 U.S. 830, 225 USPQ 232 (1984), the Federal Circuit held that, where the only difference between the prior art and the claims was a recitation of relative dimensions of the claimed device and a device having the claimed relative dimensions would not perform differently than the prior art device, the claimed device was not patentably distinct from the prior art device. In *re Dailey*, 357 F.2d 669, 149 USPQ 47 (CCPA 1966) (The court held that the configuration of the claimed disposable plastic nursing container was a matter of choice which a person of ordinary skill in the art would have found obvious absent persuasive evidence that the particular configuration of the claimed container was significant.).

Response to Arguments

Applicant's arguments filed 8/20/08 have been fully considered but they are not persuasive.

The cited references all teach sealing material that extend uniformly into the permeate/filtrate passageways to form uniform flow channels, which is clear from the cited figures.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Krishnan S. Menon whose telephone number is 571-272-1143. The examiner can normally be reached on 8:00-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, David R. Sample can be reached on 571-272-1376. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only.

Application/Control Number:
10/511,695
Art Unit: 1797

Page 7

For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Krishnan S Menon/
Primary Examiner, Art Unit 1797